

Product Test Report

Product Tested: MetPak Pro

Part Number: 1723-2B-2-111

Serial Number: 15440003

Test Date: 27/10/2015

Location: Gill Instruments Ltd

GILL ensures that quality is inherent in all aspects of their activities and ensures that compliance with BS EN ISO9001: 2008 is maintained.

This report certifies that the above instrument has been tested in accordance with Gill internal procedures

Results


Test	Limits	Results
Wind Still Air Test (Zero Wind Speed)	Pass/Fail	Pass
Wind Tunnel Test (12m/s nominal)	Pass/Fail	Pass
Pressure Sensor (Comparison DPI 142)	Pass/Fail	Pass
Temperature Sensor (Comparison HC2-S (SCS certified))	Pass/Fail	Pass
Humidity Sensor (Comparison HC2-S (SCS certified))	Pass/Fail	Pass

Wind sensor generic calibration is traceable to the University of Southampton wind tunnel and Gill instrumentation is maintained in accordance with UKAS.

Comparisons for Temperature, Humidity and Pressure are done against reference UKAS traceable instruments. The reference system numbers of these instruments are listed above.

All tests have been successfully completed

On behalf of Gill Instruments Ltd



Tony Aspland-Jupe
Quality Control



gillinstruments.com

Tel: +44 (0) 1590 613 500
anem@gillinstruments.com